

# **BigDye<sup>®</sup> Terminator v1.1 Cycle Sequencing Kit**

Protocol

© Copyright 2002, Applied Biosystems. All rights reserved.

**For Research Use Only. Not for use in diagnostic procedures.**

**Notice To Purchaser: Limited License**

A license under the process claims of U.S. Patents 5,332,666 and 5,821,058 or their foreign counterpart claims, has an up-front fee component and a running-royalty component. The purchase price of BigDye<sup>®</sup> Terminator v1.1 Cycle Sequencing Kit includes limited, non-transferable rights under the running-royalty component to use only this amount of the product to practice the DNA sequence and fragment analysis processes described in said patents when this product is used in conjunction with an Authorized DNA sequence analysis instrument whose use is covered under the up-front fee component of these patents. No other rights are granted expressly, by implication, or by estoppel, or under any other patent rights owned or licensable by Applied Biosystems. Further information relating to the purchase of licenses for DNA sequence and fragment analysis and other applications may be obtained by contacting the Director of Licensing at Applied Biosystems, 850 Lincoln Centre Drive, Foster City, CA 94404, U.S.A.

**Notice to Purchaser: Limited License**

The purchase of the BigDye<sup>®</sup> Terminator v1.1 Cycle Sequencing Kit includes a limited, nontransferable, non-exclusive license (without the right to resell, repackage, or sublicense) under the process claims of one or more of U.S. Patents 5,800,996, 5,863,727, and 5,945,526, and corresponding claims in foreign counterpart patents and patent applications, to use this product solely with an Applied Biosystems commercial automated DNA sequencing machine or other authorized automated DNA sequencing machines that have been authorized under these patents by Applied Biosystems. No license is hereby granted for the use of this kit or the reagents therein in any other automated sequencing machine. Such license is granted solely for research and other uses that are not unlawful. No other license is granted expressly, impliedly, or by estoppel. For information concerning the availability of additional licenses to practice the patented methodologies, contact: Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA. Patents are pending in countries outside the United States.

**Notice to Purchaser About Limited License**

This kit (reagent) is sold pursuant to a limited sublicense from Amersham International plc under one or more U.S. Patent Nos. 5,498,523; 4,994,372, U.S. Patent Application Serial Nos. 08/324437; 08/337615 and corresponding foreign patents and patent applications. The purchase of this kit (reagent) includes a limited non-exclusive sublicense (without the right to resell, repackage or further sublicense) under such patent rights to use this reagent for DNA sequencing or fragment length analysis solely with an Applied Biosystems commercial automated sequencing machine or other authorized DNA sequencing machines that have been authorized for such use by Applied Biosystems, or for manual DNA sequencing. No license is hereby granted for use of this kit, or the reagents therein, in any other automated sequencing machine. Such sublicense is granted solely for research or other uses that are not unlawful. No other license is granted expressly, impliedly, or by estoppel. For information concerning the availability of additional license to practice the patented methodologies, contact: Amersham Life Science, Inc., Vice President, Regulatory Affairs, P.O. Box 22400, Cleveland, Ohio 44122. Patents are pending in countries outside the United States.

ABI PRISM, Applied Biosystems, BigDye, GeneScan, MicroAmp, and Primer Island are registered trademarks of Applied Biosystems Corporation or its subsidiaries in the U.S. and certain other countries., and ABI, Avant, CATALYST, Hi-Di, POP, POP-4, POP-5, and POP-6 are trademarks of Applied Biosystems Corporation or its subsidiaries in the U.S. and certain other countries..

Centricon is a trademark of W. R. Grace and Co.

Centri-Sep is a trademark of Princeton Separations, Inc.

GeneAmp are registered trademarks of Roche Molecular Systems, Inc.

pGEM is a registered trademark of Promega Corporation.

All other trademarks are the sole property of their respective owners.

# Contents

## Chapter 1 Introduction

Chapter Summary	1-1
In This Chapter	1-1
About the Kit	1-1
New Formulation	1-1
Features and Compatibilities	1-1
BigDye Terminator v1.1 Cycle Sequencing Kit	1-2
Instruments	1-3
Instrument Platforms	1-3
Thermal Cyclers	1-3
Required Software	1-4
Dye/Filter Sets and Matrix Standards for the 310 and 377 Instruments	1-4
Dye Sets and Spectral Standards for the 3700, 3100, and 3100-Avant Instruments	1-4
Instructions	1-5
Dye Set/Primer (Mobility) Files	1-6
Reagents and Storage	1-6
Available Kits	1-6
Kit Reagents	1-6
Storage and Use of the Kits	1-7
Materials Supplied by the User	1-7
Overview	1-7
Materials for Cycle Sequencing	1-8
Materials for Purifying Extension Products	1-9
Other Equipment	1-9
Materials for Electrophoresis	1-10
General Safety	1-11

---

## Chapter 2 Preparing the Templates

Chapter Summary .....	2-1
In This Chapter .....	2-1
Control DNA Templates .....	2-1
Using Control DNA .....	2-1
Control DNA Sequence .....	2-1
An Additional Control Sold Separately .....	2-1
Template Preparation Methods .....	2-3
Single- and Double-Stranded Templates .....	2-3
BAC DNA Templates .....	2-3
PCR Templates .....	2-3
Purifying PCR Fragments .....	2-4
DNA Quality .....	2-4
Poor Template Quality .....	2-4
Contamination .....	2-5
Determining DNA Quality .....	2-5
DNA Quantity .....	2-6
Quantitating DNA .....	2-6
Template Quantity .....	2-6

## Chapter 3 Performing Cycle Sequencing

Chapter Summary .....	3-1
In This Chapter .....	3-1
Sequencing Single- and Double-Stranded DNA .....	3-2
Overview .....	3-2
Preparing the Reactions for 96-Well Reaction Plates or Microcentrifuge Tubes .....	3-2
Using BigDye Terminator Sequencing Buffer .....	3-3
Preparing the Reactions for 384-Well Plates .....	3-4
Cycle Sequencing on the System 9700, 9600, 2700, or 2400 ...	3-5
Sequencing Large DNA Templates .....	3-5
Overview .....	3-5
Thermal Cyclers .....	3-5
Preparing Sequencing Reactions .....	3-6
Performing Cycle Sequencing .....	3-7

---

## Chapter 4 Purifying Extension Products

Chapter Summary	4-1
In This Chapter	4-1
Choosing a Method of Purification	4-1
Purpose	4-1
Purification Methods	4-1
Ethanol/EDTA Precipitation	4-2
Recommended Protocol	4-2
Precipitating in 96-Well Reaction Plates	4-2
Precipitating in 384-Well Reaction Plates	4-4
Ethanol/EDTA/Sodium Acetate Precipitation	4-6
Precipitating in 96-Well Reaction Plates	4-6
Precipitating in 384-Well Reaction Plates	4-8
Plate and Spin Column Purification	4-10
Overview	4-10
Recommended Spin Columns	4-10
Optimizing Spin Column Purification	4-10
Performing Spin Column Purification	4-11
Plate Purification Using SDS/Heat Treatment	4-13
Overview	4-13
Preparing Extension Products	4-13
Performing 96-Well Spin Plate Purification	4-14
Recommended 96-Well Spin Plates	4-14
Purifying with the 96-Well Spin Plate	4-14

## Appendix A Selecting Sequencing Primers

Selecting Sequencing Primers	A-1
Overview	A-1
Optimizing Primer Selection	A-1

## Appendix B Control DNA Sequence

Control Sequence	B-1
Partial Sequence of pGEM-3Zf(+)	B-1

---

## Appendix C Sample Electrophoresis

Some Important Reminders . . . . .	C-1
Electrophoresis on the ABI PRISM 3700 DNA Analyzer . . . . .	C-2
Requirements . . . . .	C-2
Performing Sample Electrophoresis . . . . .	C-3
Electrophoresis on the ABI PRISM 3100 and 3100-Avant Genetic Analyzers . . . . .	C-4
Requirements . . . . .	C-4
Performing Sample Electrophoresis . . . . .	C-5
Electrophoresis on the ABI PRISM 310 Genetic Analyzer . . . . .	C-6
Requirements . . . . .	C-6
Performing Sample Electrophoresis . . . . .	C-7
Electrophoresis on the ABI PRISM 377 DNA Sequencers . . . . .	C-8
Requirements . . . . .	C-8
Using the Lane Guide Kit . . . . .	C-9
Using Long-Read Gel and Buffer Formulations . . . . .	C-9
Performing Sample Electrophoresis . . . . .	C-9

## Appendix D Troubleshooting

## Appendix E Obtaining Technical Support

Services and Support . . . . .	E-1
Applied Biosystems Web Site . . . . .	E-1

## Index

# Introduction

---

## Chapter Summary

**In This Chapter** The following topics are covered in this chapter:

About the Kit . . . . .	1-1
Instruments . . . . .	1-3
Required Software . . . . .	1-4
Reagents and Storage . . . . .	1-6
Materials Supplied by the User . . . . .	1-7
General Safety . . . . .	1-11

## About the Kit

**New Formulation** The BigDye® Terminator v1.1 Cycle Sequencing Kit has a new formulation that delivers:

- Increased robustness
- High quality data
- Improved success with difficult templates

The protocols for cycle sequencing have been modified to optimize results using the new formulation.

- Features and Compatibilities**
- The BigDye Terminator v1.1 Cycle Sequencing Kit uses v1.0 matrix and spectral files and does not require new instrument (matrix) files for the ABI PRISM® 310 Genetic Analyzer, and ABI PRISM® 377 DNA Sequencers or new spectral calibrations for the ABI PRISM® 3700 DNA Analyzer, ABI PRISM® 3100 Genetic Analyzer, and ABI PRISM® 3100-Avant Genetic Analyzer.

Applied Biosystems recommends that you verify the quality of your current matrix before proceeding. If it is necessary to generate a new matrix, use the appropriate matrix and/or sequencing standard for your instrument.

- The 310 and 377 instruments use the 310/377 BigDye® Terminator v1.1 Matrix Standards (PN 4336805) for instrument (matrix) file generation.
- The 3700 instrument requires the 3700 BigDye® Terminator v1.1 Matrix Standard (PN 4336825) for spectral calibration.
- The 3100 and 3100-Avant instruments require the 3100 BigDye® Terminator v1.1 Matrix Standard (PN 4336824) for spectral calibration.
- The alcohol precipitation methods are different from those recommended for the original and other versions of BigDye® terminators.
- The existing mobility files can be used with their respective platforms. New mobility files are not necessary.

### **BigDye Terminator v1.1 Cycle Sequencing Kit**

The BigDye Terminator v1.1 Cycle Sequencing Kit provides the required reagent components for the sequencing reaction in a ready reaction, pre-mixed format. You need only provide your template and template-specific primer.

These reagents are suitable for performing fluorescence-based cycle sequencing reactions on single-stranded or double-stranded DNA templates, on polymerase chain reaction (PCR) fragments, and on large templates (for example, BAC clones).

**Note:** This kit includes BigDye® Terminator v1.1/3.1 Sequencing Buffer (5X), which has been specifically optimized for use with the new BigDye® ready reaction mixes. This buffer should be used for any reaction optimization you undertake.

---

# Instruments

## Instrument Platforms

The BigDye Terminator v1.1 Cycle Sequencing Kit is for use with the following instruments:

- ABI PRISM® 3700 DNA Analyzer
- ABI PRISM® 3100 Genetic Analyzer
- ABI PRISM® 3100-Avant Genetic Analyzer
- ABI PRISM® 310 Genetic Analyzer
- ABI PRISM® 377 DNA Sequencer (all models\*)

General instructions are given for using the kit reagents to generate samples for these instruments. For more detailed instructions, refer to the appropriate instrument user's manual or chemistry guide.

## Thermal Cyclers

The protocols provided in this document were optimized using Applied Biosystems thermal cyclers, including the GeneAmp® PCR Systems 9700, 9600, 2700, and 2400.

**Note:** If you use a thermal cycler not manufactured by Applied Biosystems, you may need to optimize thermal cycling conditions. Ramping time is very important. If the thermal ramping time is too fast ( $>1^{\circ}/\text{sec}$ ), poor (noisy) data may result.

\*Includes the ABI PRISM 377, ABI PRISM 377-18, ABI PRISM 377 with XL Upgrade, and the ABI PRISM 377 with 96-Lane Upgrade instruments.

## Required Software

### Dye/Filter Sets and Matrix Standards for the 310 and 377 Instruments

The dye/filter sets and matrix standards required for the 310 and 377 instruments are listed in the table below.

Instrument	Dye/Filter Set	Standards for Instrument (Matrix) File Generation
310 Genetic Analyzer	Filter Set E	310/377 BigDye Terminator v1.1 Matrix Standards (PN 4336805)*
377 DNA Sequencer†	Filter Set E	

\*The BigDye Terminator v1.1 Cycle Sequencing Kit does not require a new spectral calibration file if you currently have a BigDye® Terminator v1.0 spectral file on your instrument.

†Includes the ABI PRISM 377, ABI PRISM 377-18, ABI PRISM 377 with XL Upgrade, and the ABI PRISM 377 with 96-Lane Upgrade instruments.

### Dye Sets and Spectral Standards for the 3700, 3100, and 3100-Avant Instruments

**IMPORTANT!** Spectral calibrations for the BigDye terminators v1.1 are not compatible with the v3.0 or v3.1 terminators.

Instrument	Dye/Filter Set	Standards for Instrument (Matrix) File Generation
3700 DNA Analyzer	Filter Set E	3700 BigDye Terminator v1.1 Matrix Standard (PN 4336825)*
3100 Genetic Analyzer	Filter Set E	3100 BigDye® Terminator v1.1 Matrix Standard (PN 4336824)*
3100-Avant Genetic Analyzer		

\*The BigDye Terminator v1.1 Cycle Sequencing Kit does not require a new spectral calibration file if you currently have a BigDye Terminator v1.0 spectral file on your instrument.

**IMPORTANT!** Users of the ABI Prism 3700 DNA Analyzer refer to the *ABI Prism 3700 DNA Analyzer User's Manual* (PN 4306152) for information on instrument (matrix) files.

## **Instructions For Generating Matrices**

For the 377 and 310 instruments, refer to the product insert (included with matrix or sequence standards) for instructions on using the BigDye Matrix Standards v1.1 to generate matrices.

## **For Performing Spectral Calibrations**

- For the 3700 instrument, refer to the product insert for instructions on using the 3700/3730 BigDye Terminator v1.1 Matrix Standard to perform spectral calibration.
- For the 3100 and 3100-Avant instruments, refer to the product insert for instructions on using the 3100 BigDye Terminator v1.1 Matrix Standard to perform spectral calibration.

**Dye Set/Primer (Mobility) Files**

To analyze sequencing data generated with BigDye® chemistries v1.1, you need dye set/primer (mobility) files that were created for v1.0 chemistries. The dye set/primer (mobility) files can be downloaded from the Internet.

Dye set/primer (mobility) files can be downloaded from the Applied Biosystems website:

**<http://www.appliedbiosystems.com/support/software>**

If you do not have access to the Internet, contact Applied Biosystems Technical Support, or your local field applications specialist (call your local sales office for more information).

## Reagents and Storage

**Available Kits** The following kits are available:

Kit	Number of Reactions	Part Number
BigDye Terminator v1.1 Cycle Sequencing Kit	100	4337450
	1000	4337451
	5000	4337452
	25,000	4337453

The *BigDye® Terminator v1.1 Cycle Sequencing Kit Protocol* (PN 4337036) is available separately and can be ordered at no charge.

**Kit Reagents**

A listing of the kit components is given below.

- Ready Reaction Mix
- pGEM®-3Zf(+) double-stranded DNA Control Template
- -21 M13 Control Primer (forward)
- BigDye Terminator v1.1/3.1 Sequencing Buffer (5X)

## Storage and Use of the Kits

- Store the kit at  $-15$  to  $-25$  °C.  
**Note:** The BigDye sequencing buffer can be stored at 4 °C.
- Avoid excess (that is, no more than 5 to 10) freeze-thaw cycles. Aliquot reagents in smaller amounts if necessary.
- Before each use of the kit, allow the frozen stocks to thaw at room temperature (do not heat).  
**IMPORTANT!** Mix each stock thoroughly and then centrifuge briefly to collect all the liquid at the bottom of each tube.
- Whenever possible, keep thawed materials on ice during use. Do not leave reagents at room temperature for extended periods.

## Materials Supplied by the User

**Overview** In addition to the reagents supplied in this kit, other items are required.

This section lists general materials needed for:

- Cycle sequencing
- Purifying extension products

**Note:** Many of the items listed in this section are available from major laboratory suppliers (MLS) unless otherwise noted. Equivalent sources may be acceptable where noted.

Refer to the individual instrument protocols for the specific items needed for each instrument.



**WARNING CHEMICAL HAZARD.** Before handling the chemical reagents needed for cycle sequencing, read the safety warnings on the reagent bottles and in the manufacturers' Material Safety Data Sheets (MSDSs), and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves. Dispose of waste in accordance with all local, state/provincial, and national environmental and health regulations.

## Materials for Cycle Sequencing

The table below lists the plates or tubes required for the recommended Applied Biosystems thermal cyclers.

Thermal Cycler	Plate or Tube	Applied Biosystems Part Number
GeneAmp® PCR System 9700	MicroAmp® 384-Well Reaction Plate	4305505
	MicroAmp® 96-Well Reaction Plate	N801-0560
	MicroAmp® Reaction Tubes, 0.2-µL	N801-0533
	MicroAmp® Caps, 12 or 8/strip	N801-0534 or N801-0535
	ABI PRISM™ Optical Adhesive Cover Starter Pack or ABI PRISM® Optical Adhesive Covers	4313663 or 4311971
GeneAmp® PCR System 9600	MicroAmp® 96-Well Reaction Plate	N801-0560
	MicroAmp® Reaction Tubes, 0.2-µL	N801-0533
	MicroAmp® Caps, 12 or 8/strip	N801-0534 N801-0535
	ABI PRISM™ Optical Adhesive Cover Starter Pack or ABI PRISM® Optical Adhesive Covers	4313663 or 4311971
GeneAmp® PCR System 2400 and 2700	MicroAmp® Reaction Tubes, 0.2-mL	N801-0533
	MicroAmp® Caps, 12 or 8/strip	N801-0534 N801-0535

## Materials for Purifying Extension Products

Method	Material	Supplier
Ethanol/EDTA Precipitation	Ethanol (EtOH), 200 proof, Molecular Biology grade EDTA, 125 mM Sealing taped	MLS  MLS Costar 6570 Thermowell Sealing Tape
Ethanol/Sodium Acetate Precipitation	Ethanol (EtOH), 200 proof, Molecular Biology grade Sodium acetate (NaOAc), 3 M, pH 5.2 Sealing tape	MLS  Applied Biosystems (PN 400320) Costar 6570 Thermowell Sealing Tape
Plate Column Purification <b>Note:</b> For 96-well reaction plates	96-Well columns for purification  Sealing tape	See "Recommended 96-Well Spin Plates" on page 4-14  Costar 6570 Thermowell Sealing Tape
Spin Column Purification	Centri-Sep™ spin column, 1-mL, 32 columns, 100 columns Sealing tape	Applied Biosystems PN 401763, PN 401762  Costar 6570 Thermowell Sealing Tape

## Other Equipment

You will also need a variable speed centrifuge with microtiter plate holders capable of reach a spin speed of at least  $1400 \times g$ . Applied Biosystems recommends a Beckman Allegra 6A centrifuge with a GH-3.8A rotor.

## Materials for Electrophoresis

Instrument	Material	Supplier
ABI PRISM 3700 DNA Analyzer	Hi-Di™ Formamide, 25-mL bottle	Applied Biosystems (PN 4311320)
	3700 BigDye Terminator v1.1 Matrix Standard	Applied Biosystems (PN 4336825)
	3700/3730 BigDye® Terminator v1.1 Sequencing Standard	Applied Biosystems (PN 4336799)
ABI PRISM 3100 and 3100-Avant Genetic Analyzers	Hi-Di™ Formamide, 25-mL bottle	Applied Biosystems (PN 4311320)
	3100 BigDye Terminator v1.1 Matrix Standard	Applied Biosystems (PN 4336824)
	BigDye® Terminator v1.1 Sequencing Standard	Applied Biosystems (PN 4336791)
ABI PRISM 310 Genetic Analyzer	Template Suppression Reagent (TSR)	Applied Biosystems (PN 401674)
	EDTA	MLS
	310/377 BigDye Terminator v1.1 Matrix Standards	Applied Biosystems (PN 4336805)
ABI PRISM 377 DNA Sequencer	Formamide	MLS
	EDTA	MLS
	25 mM EDTA with 50 mg/mL blue dextran	Applied Biosystems (PN 402055)
	310/377 BigDye Terminator v1.1 Matrix Standards	Applied Biosystems (PN 4336805)

## General Safety

### Documentation User Attention Words

Five user attention words appear in the text of all Applied Biosystems user documentation. Each word implies a particular level of observation or action as described below.

**Note:** Calls attention to useful information.

**IMPORTANT!** Indicates information that is necessary for proper instrument operation, accurate chemistry kit use, or safe use of a chemical.



**CAUTION** Indicates a potentially hazardous situation which, if not avoided, may result in minor or moderate injury. It may also be used to alert against unsafe practices.



**WARNING** Indicates a potentially hazardous situation which, if not avoided, could result in death or serious injury.



**DANGER** Indicates an imminently hazardous situation which, if not avoided, will result in death or serious injury. This signal word is to be limited to the most extreme situations.

### Site Preparation and Safety Guide

A site preparation and safety guide is a separate document sent to all customers who have purchased an Applied Biosystems instrument. Refer to the guide written for your instrument for information on site preparation, instrument safety, chemical safety, and waste profiles.

## Chemical Safety

### Chemical Hazard Warning



**WARNING CHEMICAL HAZARD.** Some of the chemicals used with Applied Biosystems instruments and protocols are potentially hazardous and can cause injury, illness, or death.

- Read and understand the material safety data sheets (MSDSs) provided by the chemical manufacturer before you store, handle, or work with any chemicals or hazardous materials.
- Minimize contact with chemicals. Wear appropriate personal protective equipment when handling chemicals (*e.g.*, safety glasses, gloves, or protective clothing). For additional safety guidelines, consult the MSDS.

- Minimize the inhalation of chemicals. Do not leave chemical containers open. Use only with adequate ventilation (*e.g.*, fume hood). For additional safety guidelines, consult the MSDS.
- Check regularly for chemical leaks or spills. If a leak or spill occurs, follow the manufacturer's cleanup procedures as recommended on the MSDS.
- Comply with all local, state/provincial, or national laws and regulations related to chemical storage, handling, and disposal.

## Chemical Waste Hazard Warning



**WARNING CHEMICAL WASTE HAZARD.** Wastes produced by Applied Biosystems instruments are potentially hazardous and can cause injury, illness, or death.

- Read and understand the material safety data sheets (MSDSs) provided by the manufacturers of the chemicals in the waste container before you store, handle, or dispose of chemical waste.
- Handle chemical wastes in a fume hood.
- Minimize contact with chemicals. Wear appropriate personal protective equipment when handling chemicals (*e.g.*, safety glasses, gloves, or protective clothing). For additional safety guidelines, consult the MSDS.
- Minimize the inhalation of chemicals. Do not leave chemical containers open. Use only with adequate ventilation (*e.g.*, fume hood). For additional safety guidelines, consult the MSDS.
- After emptying the waste container, seal it with the cap provided.
- Dispose of the contents of the waste tray and waste bottle in accordance with good laboratory practices and local, state/provincial, or national environmental and health regulations.

## About MSDSs

Some of the chemicals used with this instrument may be listed as hazardous by their manufacturer. When hazards exist, warnings are prominently displayed on the labels of all chemicals.

Chemical manufacturers supply a current material safety data sheet (MSDS) before or with shipments of hazardous chemicals to new customers and with the first shipment of a hazardous chemical after an MSDS update. MSDSs provide you with the safety information you need to store, handle, transport and dispose of the chemicals safely.

We strongly recommend that you replace the appropriate MSDS in your files each time you receive a new MSDS packaged with a hazardous chemical.



**WARNING CHEMICAL HAZARD.** Be sure to familiarize yourself with the MSDSs before using reagents or solvents.

## Ordering MSDSs

You can order free additional copies of MSDSs for chemicals manufactured or distributed by Applied Biosystems using the contact information below.

To order documents by automated telephone service:

1.	From the U.S. or Canada, dial <b>1.800.487.6809</b> .
2.	Follow the voice instructions to order documents (for delivery by fax).  <b>Note:</b> There is a limit of five documents per fax request.

To obtain documents through the Applied Biosystems Web site:

1.	Go to <b><a href="http://docs.appliedbiosystems.com/msdssearch.html">http://docs.appliedbiosystems.com/msdssearch.html</a></b>
2.	In the <b>SEARCH</b> field, type in the chemical name, part number, or other information that will appear in the MSDS and click <b>SEARCH</b> .  <b>Note:</b> You may also select the language of your choice from the drop-down list.
3.	When the <b>Search Results</b> page opens, find the document you want and click on it to open a PDF of the document.

For chemicals not manufactured or distributed by Applied Biosystems, call the chemical manufacturer.



# Preparing the Templates

---

## Chapter Summary

**In This Chapter** The following topics are covered in this chapter:

Control DNA Templates . . . . .	2-1
Template Preparation Methods . . . . .	2-3
DNA Quality . . . . .	2-4
DNA Quantity . . . . .	2-6

## Control DNA Templates

**Using Control DNA** Include a control DNA template as one of the templates in a set of sequencing reactions. The results from the control can help determine whether failed reactions are the result of poor template quality or sequencing reaction failure.

**Control DNA Sequence** Applied Biosystems recommends M13mp18 as a single-stranded control and pGEM<sup>®</sup>-3Zf(+) as a double-stranded control. All Applied Biosystems DNA sequencing kits provide pGEM<sup>®</sup> control DNA. All dye terminator cycle sequencing kits include a –21 M13 forward primer for use in performing control reactions.

The partial sequence of pGEM-3Zf(+) from the –21 M13 forward primer, followed by the ensuing 1000 bases is shown in Appendix A, “Control DNA Sequence.”

**An Additional Control Sold Separately** The BigDye<sup>®</sup> Terminator v1.1 Sequencing Standard provides an additional control to help in troubleshooting electrophoresis runs. This standard contains lyophilized sequencing reactions that require only resuspension and denaturation before use.

There are two forms of the v1.0 and v1.1 sequencing standard, as shown in the table below. Please use the correct sequencing standard for your instrument. Refer to the product inserts for instructions on using each sequencing standard.

<b>Instrument</b>	<b>Kit</b>	<b>PN</b>
ABI PRISM® 3700 DNA Analyzer	3700/3730 BigDye® Terminator v1.1 Sequencing Standard	4336799
ABI PRISM® 3100 Genetic Analyzer	BigDye® Terminator v1.1 Sequencing Standard	4336791
ABI PRISM® 3100-Avant Genetic Analyzer		
ABI PRISM® 310 Genetic Analyzer		
ABI PRISM® 377 DNA Sequencers*		

\*Includes the ABI PRISM 377, ABI PRISM 377-18, ABI PRISM 377 with XL Upgrade, and the ABI PRISM 377 with 96-Lane Upgrade instruments.

# Template Preparation Methods

## Single- and Double-Stranded Templates

Refer to the *Automated DNA Sequencing Chemistry Guide* (PN 4305080) for information on preparing single- and double-stranded templates.

## BAC DNA Templates

With larger DNA targets such as bacterial artificial chromosomes (BACs), the quality of DNA template is important to the success of the sequencing reaction. Two methods have given good sequencing results:

- Alkaline lysis,\* with extra phenol extraction and isopropanol precipitation if very clean DNA is desired
- Cesium chloride (CsCl) banding

### Commercial Kits

Commercial kits are also available for BAC DNA preparation:

- QIAGEN-tip 100 (QIAGEN: PN 10043, 25 reactions; 10045, 100 reactions)
- QIAGEN-tip 500 (QIAGEN: PN 10063, 25 reactions; 10065, 100 reactions)

## PCR Templates

Cycle sequencing provides the most reproducible results for sequencing PCR templates. Although PCR fragments can be difficult to denature with traditional sequencing methods, cycle sequencing provides several chances to denature and extend the template, ensuring adequate signal in the sequencing reaction.

### Importance of Purifying Product

For optimum results, purify the PCR product before sequencing. In general, any method that removes dNTPs and primers should work. We recommend Centricon®-100 columns (PN N930-2119). The protocol for using these columns is provided in “Purifying PCR Fragments” on page 2-4

Refer to the *Automated DNA Sequencing Chemistry Guide* for information on sequencing PCR templates.

\*Marra, M., Weinstock, L.A., and Mardis, E.R. 1966. End sequence determination from large insert cloning using energy transfer fluorescent primers. *Genomic Methods* 6: 1118-1122.

## Purifying PCR Fragments

To purify PCR fragments by ultrafiltration:

1.	Assemble the Centricon-100 column according to the manufacturer's recommendations.
2.	Load 2 mL deionized water onto the column.
3.	Add the entire sample to the column.
4.	Spin the column at $3000 \times g$ in a fixed-angle centrifuge for 10 minutes.  <b>Note:</b> The manufacturer recommends a maximum speed of $1000 \times g$ , but $3000 \times g$ has worked well in Applied Biosystems laboratories. If you are following the manufacturer's guidelines, increase the time to compensate.
5.	Remove the waste receptacle and attach the collection vial.
6.	Invert the column and spin it at $270 \times g$ for 2 minutes to collect approximately 40–60 $\mu\text{L}$ of sample.
7.	Add deionized water to bring the purified PCR fragments to the original volume.

## DNA Quality

### Poor Template Quality

Poor template quality is the most common cause of sequencing problems. The following are characteristics of poor quality templates:

- Noisy data or peaks under peaks
- No usable sequence data
- Weak signal

Always follow recommended procedures to prepare templates.

---

**Contamination** Potential contaminants include:

- Proteins
- RNA
- Chromosomal DNA
- Excess PCR primers, dNTPs, enzyme, and buffer components (from a PCR amplification used to generate the sequencing template)
- Residual salts
- Residual organic chemicals such as phenol, chloroform, and ethanol
- Residual detergents

**Determining DNA Quality** The following methods can be used to examine DNA quality:

- Agarose gel electrophoresis  
Purified DNA should run as a single band on an agarose gel.

**Note:** Uncut plasmid DNA can run as three bands: supercoiled, nicked, and linear.

- Spectrophotometry  
The  $A_{260}/A_{280}$  ratio should be 1.7 to 1.9. Smaller ratios usually indicate contamination by protein or organic chemicals.

Agarose gels reveal the presence of contaminating DNAs and RNAs, but not proteins. Spectrophotometry can indicate the presence of protein contamination, but not DNA and RNA contamination. Use these methods together to get the most information about your DNA before sequencing.

## DNA Quantity

**Quantitating DNA** If possible, quantitate the amount of purified DNA by measuring the absorbance at 260 nm or by some other method.

**Template Quantity** The table below shows the amount of template to use in a cycle sequencing reaction.

Template	Quantity
PCR product:	
100–200 bp	1–3 ng
200–500 bp	3–10 ng
500–1000 bp	5–20 ng
1000–2000 bp	10–40 ng
>2000 bp	20–50 ng
Single-stranded	25–50 ng
Double-stranded	150–300 ng
Cosmid, BAC	0.5–1.0 $\mu$ g
Bacterial genomic DNA	2–3 $\mu$ g

**Note:** In general, higher DNA quantities give higher signal intensities. Higher DNA quantities may also give shorter read lengths and top-heavy data.

The template quantities given above should work with all primers. You may be able to use even less DNA when using capillary instruments for detection. The amount of PCR product to use in sequencing also depends on the length and purity of the PCR product.

# Performing Cycle Sequencing

---

# 3

## Chapter Summary

**In This Chapter** The following topics are covered in this chapter:

Sequencing Single- and Double-Stranded DNA . . . . .	3-2
Sequencing Large DNA Templates . . . . .	3-5

## Sequencing Single- and Double-Stranded DNA

**Overview** This section describes how to prepare reactions and perform cycle sequencing on a variety of templates, including M13, plasmids, and PCR products.

### Preparing the Reactions for 96-Well Reaction Plates or Microcentrifuge Tubes

The type of tube required depends on the thermal cycler that you are using. Refer to “Materials for Cycle Sequencing” on page 1-8.

To prepare the reaction mixtures:

1.	For each reaction add the following reagents to a separate tube:	
	<b>Reagent</b>	<b>Quantity</b>
	Terminator Ready Reaction Mix*	8.0 $\mu$ L
	Template	See the table in “Template Quantity” on page 2-6.
	Primer	3.2 pmol
	Deionized water	q.s.
	Total Volume	20 $\mu$ L
2.	Mix well and spin briefly.	

\*See “Using BigDye Terminator Sequencing Buffer” below.

## Using BigDye Terminator Sequencing Buffer

The BigDye<sup>®</sup> Terminator v1.1/3.1 Sequencing Buffer (5X)\* is supplied at a 5X concentration. If you use it for sequencing reactions, be sure the final reaction volume is at a concentration of 1X. For example, for a half reaction in 20 µL final volume, you would use 4 µL of ready reaction premix and 2 µL of BigDye sequencing buffer as shown below.

Reagent	Concentration	Volume
Ready Reaction Premix	2.5X	4 µL
BigDye Sequencing Buffer	5X	2 µL
Primer	—	3.2 pmol
Template	—	See “Template Quantity” on page 2-6.
Water	—	to 20 µL
<b>Final Volume</b>	<b>1X</b>	<b>20 µL</b>

**Note:** The use of this buffer without optimization may result in deterioration of sequence quality. Applied Biosystems does not support diluted reactions or guarantee the performance of BigDye<sup>®</sup> chemistry when it is diluted.

\*The BigDye Terminator v1.1/3.1 Sequencing Buffer is intended for use only with BigDye Terminator v1.1/3.1 Cycle Sequencing Kits.

## Preparing the Reactions for 384-Well Plates

The type of tube required depends on the thermal cycler that you are using. Refer to “Materials for Cycle Sequencing” on page 1-8.

**Note:** The wells in a 384-well reaction plate have a volume capacity of 35  $\mu\text{L}$ . Therefore, Applied Biosystems recommends doing a 10  $\mu\text{L}$  reaction. This allows the post-reaction cleanup step to be performed in the same well.

To prepare the reaction mixtures:

1.	For each reaction add the following reagents to a separate tube:	
	<b>Reagent</b>	<b>Quantity</b>
	Terminator Ready Reaction Mix*	4.0 $\mu\text{L}$
	Template	See the table in “Template Quantity” on page 2-6.
	Primer	3.2 pmol
	Deionized water	q.s.
	Total Volume	10 $\mu\text{L}$
2.	Mix well and spin briefly.	
3.	Use on a GeneAmp® PCR System 9700 Dual 384-Well Sample Block Module.	

\*Note: For instructions on using BigDye sequencing buffer, see “Using BigDye Terminator Sequencing Buffer” on page 3-3.

**Cycle  
Sequencing on  
the System 9700,  
9600, 2700, or  
2400**

To sequence single- and double-stranded DNA on the GeneAmp® PCR System 9700 (in 9600 emulation mode), 9600, 2700, or 2400:

1.	Place the tubes in a thermal cycler and set to the correct volume.
2.	Perform an initial denaturation. <ol style="list-style-type: none"> <li>a. Rapid thermal ramp to 96 °C</li> <li>b. 96 °C for 1 min</li> </ol>
3.	Repeat the following for 25 cycles: <ul style="list-style-type: none"> <li>• Rapid thermal ramp* to 96 °C</li> <li>• 96 °C for 10 seconds.</li> <li>• Rapid thermal ramp to 50 °C</li> <li>• 50 °C for 5 seconds.</li> <li>• Rapid thermal ramp to 60 °C</li> <li>• 60 °C for 4 minutes</li> </ul>
4.	Rapid thermal ramp to 4 °C and hold until ready to purify.
5.	Spin down the contents of the tubes in a microcentrifuge.
6.	Proceed to Chapter 4, “Purifying Extension Products.”

\*Rapid thermal ramp is 1 °C/second.

## Sequencing Large DNA Templates

**Overview** This section describes how to prepare reactions and perform cycle sequencing on large DNA templates such as:

- BAC DNA
- Cosmid DNA
- Genomic DNA

**Thermal Cyclers** Only the following thermal cyclers can be used with this protocol:

- GeneAmp PCR System 9600
- GeneAmp PCR System 9700 (in 9600 emulation mode)

Reoptimize this protocol for use on other thermal cyclers.

## Preparing Sequencing Reactions

The type of tube required depends on the thermal cycler that you are using. Refer to “Materials Supplied by the User” on page 1-7.

To prepare the sequencing reaction:

1.	For each reaction, add the following reagents to a separate tube:												
<table border="1"> <thead> <tr> <th data-bbox="491 479 928 539">Reagent</th> <th data-bbox="935 479 1173 539">Quantity</th> </tr> </thead> <tbody> <tr> <td data-bbox="491 548 928 591">Terminator Ready Reaction Mix*</td> <td data-bbox="935 548 1173 591">8.0 <math>\mu</math>L</td> </tr> <tr> <td data-bbox="491 600 928 704">DNA Template</td> <td data-bbox="935 600 1173 704">See the table in “DNA Quantity” on page 2-6</td> </tr> <tr> <td data-bbox="491 713 928 756">Primer</td> <td data-bbox="935 713 1173 756">3.2 pmol</td> </tr> <tr> <td data-bbox="491 765 928 808">Deionized water</td> <td data-bbox="935 765 1173 808">q.s.</td> </tr> <tr> <td data-bbox="491 817 928 861">Total Volume</td> <td data-bbox="935 817 1173 861">20 <math>\mu</math>L</td> </tr> </tbody> </table>		Reagent	Quantity	Terminator Ready Reaction Mix*	8.0 $\mu$ L	DNA Template	See the table in “DNA Quantity” on page 2-6	Primer	3.2 pmol	Deionized water	q.s.	Total Volume	20 $\mu$ L
Reagent	Quantity												
Terminator Ready Reaction Mix*	8.0 $\mu$ L												
DNA Template	See the table in “DNA Quantity” on page 2-6												
Primer	3.2 pmol												
Deionized water	q.s.												
Total Volume	20 $\mu$ L												
2.	Mix well and spin briefly.												

\*Note: For instructions on using BigDye sequencing buffer, see “Using BigDye Terminator Sequencing Buffer” on page 3-3.

**Performing Cycle Sequencing****To perform cycle sequencing on BAC DNA:**

1.	Place the tubes in a thermal cycler and set the volume to 20 $\mu$ L.
2.	Heat the tubes at 95 °C for 5 minutes.
3.	Repeat the following for 30 cycles:* <ul style="list-style-type: none"><li>• Rapid thermal ramp† to 95 °C</li><li>• 95 °C for 30 seconds.</li><li>• Rapid thermal ramp to 50–55 °C (depending on template)</li><li>• 50–55 °C for 10 seconds.</li><li>• Rapid thermal ramp to 60 °C</li><li>• 60 °C for 4 minutes.</li></ul>
4.	Rapid thermal ramp to 4 °C and hold until ready to purify.
5.	Spin down the contents of the tubes in a microcentrifuge.
6.	Proceed to Chapter 4, “Purifying Extension Products.”

\*Some laboratories have found that increasing the number of cycles gives better results.

†Rapid thermal ramp is 1 °C/sec.



# Purifying Extension Products

---

## Chapter Summary

**In This Chapter** The following topics are covered in this chapter:

Choosing a Method of Purification . . . . .	4-1
Ethanol/EDTA Precipitation . . . . .	4-2
Ethanol/EDTA/Sodium Acetate Precipitation . . . . .	4-6
Plate and Spin Column Purification . . . . .	4-10
Plate Purification Using SDS/Heat Treatment . . . . .	4-13
Performing 96-Well Spin Plate Purification . . . . .	4-14

## Choosing a Method of Purification

**Purpose** The best results are obtained when unincorporated dye terminators are completely removed prior to electrophoresis. Excess dye terminators in sequencing reactions obscure data in the early part of the sequence and can interfere with basecalling.

**Purification Methods** The methods recommended below have produced clean sequencing data for BigDye<sup>®</sup> Terminator v1.1 Cycle Sequencing Kit products:

- Ethanol/EDTA precipitation
- Ethanol/EDTA/sodium acetate precipitation
- Plate and Spin column purification

Earlier methods developed for other sequencing kits may not remove unincorporated dyes efficiently.

Use the method that works best for your particular application.

**Note:** These precipitation protocols have been optimized for use with the v1.1 formulation at the specified sequencing reaction volumes (20  $\mu\text{L}$  in 96-well format and 10  $\mu\text{L}$  in 384-well format) and are not recommended for other versions.

**IMPORTANT!** To clean up sequencing reactions at volumes less than those specified, reduce each component of the precipitation protocol proportionately.

## Ethanol/EDTA Precipitation

### Recommended Protocol

With the BigDye<sup>®</sup> terminators v1.1, the ethanol/EDTA precipitation method produces consistent signal, while minimizing unincorporated dyes. It is particularly good at getting rid of unincorporated dye-labelled terminators that obscure data in the early part of the sequence.

**Note:** While this method produces the cleanest signal, it may cause loss of small molecular weight fragments.

**IMPORTANT!** Absolute ethanol absorbs water from the atmosphere, gradually decreasing its concentration and leading to inaccurate final concentrations of ethanol, which can affect some sequencing results.

**IMPORTANT!** 95% ethanol is usable, but you must make sure the final ethanol concentration for precipitation remains the same (67–71%).

### Precipitating in 96-Well Reaction Plates



**WARNING CHEMICAL HAZARD. EDTA.** Exposure causes eye irritation. Read the MSDS, and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves.



**WARNING CHEMICAL HAZARD. Ethanol** is a flammable liquid and vapor. Exposure causes eye, skin, and respiratory tract irritation and may cause central nervous system depression and liver damage. Read the MSDS, and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves.

To precipitate 20  $\mu\text{L}$  sequencing reactions in 96-well reaction plates:

1.	Remove the 96-well reaction plate from the thermal cycler and briefly spin.						
2.	Add 5 $\mu\text{L}$ of 125 mM EDTA to each well. <b>Note:</b> Make sure the EDTA reaches the bottom of the wells.						
3.	Add 60 $\mu\text{L}$ of 100% ethanol to each well.						
4.	Seal the plate with aluminum tape and mix by inverting 4 times.						
5.	Incubate at room temperature for 15 min.						
6.	<table border="1"> <thead> <tr> <th>If you are using ...</th> <th>Then ...</th> </tr> </thead> <tbody> <tr> <td>a Beckman Allegra 6A centrifuge with a GH-3.8A rotor</td> <td>set it at 4 °C and spin the plate at 1650 <math>\times g</math> for 45 min.</td> </tr> <tr> <td>any other centrifuge</td> <td>use a plate adapter and spin the plate at the maximum speed as follows: <ul style="list-style-type: none"> <li>• 1400–2000 <math>\times g</math> for 45 min</li> <li>or</li> <li>• 2000–3000 <math>\times g</math> for 30 min</li> </ul> </td> </tr> </tbody> </table> <p><b>IMPORTANT!</b> Proceed to the next step immediately. If this is not possible, then spin the plate for an additional 2 min before performing the next step.</p>	If you are using ...	Then ...	a Beckman Allegra 6A centrifuge with a GH-3.8A rotor	set it at 4 °C and spin the plate at 1650 $\times g$ for 45 min.	any other centrifuge	use a plate adapter and spin the plate at the maximum speed as follows: <ul style="list-style-type: none"> <li>• 1400–2000 <math>\times g</math> for 45 min</li> <li>or</li> <li>• 2000–3000 <math>\times g</math> for 30 min</li> </ul>
If you are using ...	Then ...						
a Beckman Allegra 6A centrifuge with a GH-3.8A rotor	set it at 4 °C and spin the plate at 1650 $\times g$ for 45 min.						
any other centrifuge	use a plate adapter and spin the plate at the maximum speed as follows: <ul style="list-style-type: none"> <li>• 1400–2000 <math>\times g</math> for 45 min</li> <li>or</li> <li>• 2000–3000 <math>\times g</math> for 30 min</li> </ul>						
7.	Invert the plate and spin up to 185 $\times g$ , then remove from the centrifuge.						
8.	Add 60 $\mu\text{L}$ of 70% ethanol to each well.						
9.	With the centrifuge set to 4 °C, spin at 1650 $\times g$ for 15 min.						
10.	Invert the plate and spin up to 185 $\times g$ for 1 min, then remove from the centrifuge. <b>Note:</b> Start timing when the rotor starts moving.						

To precipitate 20  $\mu$ L sequencing reactions in 96-well reaction plates: (continued)

- |     |  |
|-----|--|
| 11. | <p>To continue, resuspend the samples in injection buffer. To store, cover with aluminum foil, and store at 4 °C.</p> <p><b>Note:</b> Make sure the wells are dry. You may use a Speed-Vac for 15 min to dry the plate.</p> <p><b>IMPORTANT!</b> Make sure the samples are protected from light while they are drying.</p> |
|-----|--|

Precipitating in  
384-Well  
Reaction Plates



**WARNING CHEMICAL HAZARD. EDTA.** Exposure causes eye irritation. Read the MSDS, and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves.



**WARNING CHEMICAL HAZARD. Ethanol** is a flammable liquid and vapor. Exposure causes eye, skin, and respiratory tract irritation and may cause central nervous system depression and liver damage. Read the MSDS, and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves.

To precipitate in 384-well reaction plates:

1.	Remove the 384-well reaction plates from the thermal cycler. Remove the seal from each plate and briefly spin the plates.
2.	Add 2.5 $\mu$ L of 125 mM EDTA to each well. <b>Note:</b> Make sure the EDTA reaches the bottom of the wells.
3.	Add 25 $\mu$ L of 100% ethanol to each well.
4.	Seal the plates with aluminum tape and mix by inverting 4 times.
5.	Incubate at room temperature for 15 min.

## To precipitate in 384-well reaction plates: (continued)

6.	<b>If you are using ...</b>	<b>Then ...</b>
	a Beckman Allegra 6A centrifuge with a GH-3.8A rotor	set it at 4 °C and spin the plate at 1650 × g for 45 min.
	any other centrifuge	use a plate adapter and spin the plate at the maximum speed as follows: <ul style="list-style-type: none"> <li>• 1400–2000 × g for 45 min</li> <li>or</li> <li>• 2000–3000 × g for 30 min</li> </ul>
<b>IMPORTANT!</b> Proceed to the next step immediately. If this is not possible, then spin the plate for an additional 2 min before performing the next step.		
7.	Invert the plate and spin up to 185 × g, then remove from the centrifuge.	
8.	Add 30 μL of 70% ethanol to each well.	
9.	With the centrifuge set to 4 °C, spin at 1650 × g for 15 min.	
10.	Invert the plate and spin up to 185 × g for 1 min, then remove from the centrifuge. <b>Note:</b> Start timing when the rotor starts moving.	
11.	To continue, resuspend the samples in injection buffer. To store, cover with aluminum foil, and store at 4 °C. <b>Note:</b> Make sure the wells are dry. You may use a Speed-Vac for 15 min to dry the plate. <b>IMPORTANT!</b> Make sure the samples are protected from light while they are drying.	

## Ethanol/EDTA/Sodium Acetate Precipitation

**Note:** Ethano/EDTA/sodium acetate precipitation is recommended when good signal from base 1 is required. However, for reactions containing high concentrations of unincorporated terminators, some residual terminators may be carried through the precipitation. To completely remove excess terminators in these cases, ethanol/EDTA precipitation is recommended (see page 4-2).

### Precipitating in 96-Well Reaction Plates



**WARNING CHEMICAL HAZARD. EDTA.** Exposure causes eye irritation. Read the MSDS, and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves.



**WARNING CHEMICAL HAZARD. Ethanol** is a flammable liquid and vapor. Exposure causes eye, skin, and respiratory tract irritation and may cause central nervous system depression and liver damage. Read the MSDS, and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves.

**To precipitate 20  $\mu$ L sequencing reactions in 96-well reaction plates:**

1.	Remove the 96-well reaction plate from the thermal cycler and briefly spin.
2.	Add 2 $\mu$ L of 125 mM EDTA to each well. <b>Note:</b> Make sure the EDTA reaches the bottom of the wells.
3.	Add 2 $\mu$ L of 3 M sodium acetate to each well. <b>Note:</b> Make sure the sodium acetate reaches the bottom of the wells.
4.	Add 50 $\mu$ L of 100% ethanol to each well.
5.	Seal the plate with aluminum tape and mix by inverting 4 times.
6.	Incubate at room temperature for 15 min.

To precipitate 20  $\mu\text{L}$  sequencing reactions in 96-well reaction plates: (continued)

7.	<b>If you are using ...</b>	<b>Then ...</b>
	a Beckman Allegra 6A centrifuge with a GH-3.8A rotor	set it at 4 °C and spin the plate at 1650 $\times g$ for 45 min.
	any other centrifuge	use a plate adapter and spin the plate at the maximum speed as follows: <ul style="list-style-type: none"> <li>• 1400–2000 <math>\times g</math> for 45 min</li> <li>or</li> <li>• 2000–3000 <math>\times g</math> for 30 min</li> </ul>
<p><b>IMPORTANT!</b> Proceed to the next step immediately. If this is not possible, then spin the plate for an additional 2 min before performing the next step.</p>		
8.	Invert the plate and spin up to 185 $\times g$ , then remove from the centrifuge.	
9.	Add 70 $\mu\text{L}$ of 70% ethanol to each well.	
10.	With the centrifuge set to 4 °C, spin at 1650 $\times g$ for 15 min.	
11.	<p>Invert the plate and spin up to 185 <math>\times g</math> for 1 min, then remove from the centrifuge.</p> <p><b>Note:</b> Start timing when the rotor starts moving.</p>	
12.	<p>To continue, resuspend the samples in injection buffer. To store, cover with aluminum foil, and store at 4 °C.</p> <p><b>Note:</b> Make sure the wells are dry. You may use a Speed-Vac for 15 min to dry the plate.</p> <p><b>IMPORTANT!</b> Make sure the samples are protected from light while they are drying.</p>	

## Precipitating in 384-Well Reaction Plates



**WARNING CHEMICAL HAZARD. EDTA.** Exposure causes eye irritation. Read the MSDS, and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves.



**WARNING CHEMICAL HAZARD. Ethanol** is a flammable liquid and vapor. Exposure causes eye, skin, and respiratory tract irritation and may cause central nervous system depression and liver damage. Read the MSDS, and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves.

**To precipitate 10  $\mu$ L sequencing reactions in 384-well reaction plates:**

1.	Remove the 384-well reaction plate from the thermal cycler and briefly spin.
2.	Add 1 $\mu$ L of 125 mM EDTA to each well. <b>Note:</b> Make sure the EDTA reaches the bottom of the wells.
3.	Add 1 $\mu$ L of 3 M sodium acetate to each well. <b>Note:</b> Make sure the sodium acetate reaches the bottom of the wells.
4.	Add 25 $\mu$ L of 100% ethanol to each well.
5.	Seal the plate with aluminum tape and mix by inverting 4 times.
6.	Incubate at room temperature for 15 min.

To precipitate 10  $\mu\text{L}$  sequencing reactions in 384-well reaction plates: (continued)

7.	<table border="1"> <thead> <tr> <th>If you are using ...</th> <th>Then ...</th> </tr> </thead> <tbody> <tr> <td>a Beckman Allegra 6A centrifuge with a GH-3.8A rotor</td> <td>set it at 4 °C and spin the plate at 1650 <math>\times g</math> for 45 min.</td> </tr> <tr> <td>any other centrifuge</td> <td>           use a plate adapter and spin the plate at the maximum speed as follows:           <ul style="list-style-type: none"> <li>• 1400–2000 <math>\times g</math> for 45 min</li> <li>or</li> <li>• 2000–3000 <math>\times g</math> for 30 min</li> </ul> </td> </tr> </tbody> </table>	If you are using ...	Then ...	a Beckman Allegra 6A centrifuge with a GH-3.8A rotor	set it at 4 °C and spin the plate at 1650 $\times g$ for 45 min.	any other centrifuge	use a plate adapter and spin the plate at the maximum speed as follows: <ul style="list-style-type: none"> <li>• 1400–2000 <math>\times g</math> for 45 min</li> <li>or</li> <li>• 2000–3000 <math>\times g</math> for 30 min</li> </ul>
	If you are using ...	Then ...					
	a Beckman Allegra 6A centrifuge with a GH-3.8A rotor	set it at 4 °C and spin the plate at 1650 $\times g$ for 45 min.					
any other centrifuge	use a plate adapter and spin the plate at the maximum speed as follows: <ul style="list-style-type: none"> <li>• 1400–2000 <math>\times g</math> for 45 min</li> <li>or</li> <li>• 2000–3000 <math>\times g</math> for 30 min</li> </ul>						
<p><b>IMPORTANT!</b> Proceed to the next step immediately. If this is not possible, then spin the plate for an additional 2 min before performing the next step.</p>							
8.	Invert the plate and spin up to 185 $\times g$ , then remove from the centrifuge.						
9.	Add 35 $\mu\text{L}$ of 70% ethanol to each well.						
10.	With the centrifuge set to 4 °C, spin at 1650 $\times g$ for 15 min.						
11.	<p>Invert the plate and spin up to 185 <math>\times g</math> for 1 min, then remove from the centrifuge.</p> <p><b>Note:</b> Start timing when the rotor starts moving.</p>						
12.	<p>To continue, resuspend the samples in injection buffer. To store, cover with aluminum foil, and store at 4 °C.</p> <p><b>Note:</b> Make sure the wells are dry. You may use a Speed-Vac for 15 min to dry the plate.</p> <p><b>IMPORTANT!</b> Make sure the samples are protected from light while they are drying.</p>						

## Plate and Spin Column Purification

**Overview** This section describes the recommended plate and spin columns for purifying extension products.

**IMPORTANT!** Extra caution is required when dispensing samples onto the column bed. Residual dye peaks will result if samples flow through the sides of the column.

**Recommended Spin Columns** We recommend Centri-Sep™ spin columns (Applied Biosystems, PN 401763 for 32 columns and PN 401762 for 100 columns).

**Optimizing Spin Column Purification** **IMPORTANT!** When using the BigDye terminators v1.1, hydrate the column for 2 hours (see below).

Tips for optimizing spin column purification when using individual columns:

- Do not process more columns than you can handle conveniently at one time.
- Load the sample in the center of the column bed slowly. Make sure that the sample does not touch the sides of the column and that the pipet tip does not touch the gel surface.

**Note:** If samples are not properly loaded, peaks from unincorporated dye terminators can result.

- Spin the column at 325–730 × *g* for best results. Use the following formula to calculate the best speed for your centrifuge:

$$g = 11.18 \times r \times (\text{rpm}/1000)^2$$

where:

*g* = relative centrifugal force

*r* = radius of the rotor in cm

rpm = revolutions per minute

- Do not spin for more than 2 minutes.
- Perform the entire procedure without interruption to ensure optimal results. Do not allow the column to dry out.

## Performing Spin Column Purification

To perform spin column purification:

1.	Gently tap the column to cause the gel material to settle to the bottom of the column.
2.	Remove the upper end cap and add 0.8 mL of deionized water.
3.	Replace the upper end cap and vortex or invert the column a few times to mix the water and gel material.
4.	Allow the gel to hydrate at room temperature for at least 2 hours.  <b>Note:</b> Hydrated columns can be stored for a few days at 2–6 °C. Longer storage in water is not recommended. Allow columns stored at 2–6 °C to warm to room temperature before use.
5.	Remove any air bubbles by inverting or tapping the column and allowing the gel to settle.
6.	Remove the upper end cap first, then remove the bottom cap. Allow the column to drain completely by gravity.  <b>Note:</b> If flow does not begin immediately, apply gentle pressure to the column with a pipette bulb.
7.	Insert the column into the wash tube provided.
8.	Spin the column in a microcentrifuge at 730 × g for 2 minutes to remove the interstitial fluid.
9.	Remove the column from the wash tube and insert it into a sample collection tube (for example, a 1.5-mL microcentrifuge tube).
10.	Remove the extension reaction mixture from its tube and load it carefully onto the center of the gel material.

**To perform spin column purification: (continued)**

11.	Spin the column in a microcentrifuge at $730 \times g$ for 2 minutes.  <b>Note:</b> If using a centrifuge with a fixed-angle rotor, place the column in the same orientation as it was in for the first spin. This is important because the surface of the gel will be at an angle in the column after the first spin.
12.	Discard the column. The sample is in the sample collection tube.
13.	Dry the sample in a vacuum centrifuge for 10–15 minutes, or until dry. Do not over-dry.

# Plate Purification Using SDS/Heat Treatment

**Overview** This section provides instructions for adding a sodium dodecyl sulfate (SDS)/heat treatment to the spin plate purification method. This SDS/heat treatment effectively eliminates high concentrations of unincorporated dye terminators from cycle sequencing reactions.

**Preparing Extension Products** Use this procedure to prepare extension products for 96-well spin plate purification.



**DANGER CHEMICAL HAZARD. Sodium dodecyl sulfate (SDS).** Exposure causes eye, skin, and respiratory tract irritation. Exposure may cause severe allergic respiratory and skin reaction. Read the MSDS, and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves.

**To prepare extension products:**

1.	Prepare a solution of 2.2% SDS in deionized water. This SDS solution is stable at room temperature.
2.	Add an appropriate amount of the 2.2% SDS solution to each sample to bring the final SDS concentration to 0.2%. For example: Add 2 $\mu\text{L}$ of 2.2% SDS solution to each 20- $\mu\text{L}$ completed cycle sequencing reaction.
3.	Seal the tubes and mix thoroughly.
4.	Heat the tubes to 98 °C for 5 min, then allow the tubes to cool to ambient temperature before proceeding to the next step.  <b>Note:</b> A convenient way to perform this heating/cooling cycle is to place the tubes in a thermal cycler and set it as follows: <ul style="list-style-type: none"> <li>• 98 °C for 5 min</li> <li>• 25 °C for 10 min</li> </ul>
5.	Spin down the contents of the tubes briefly.
6.	Continue with 96-well plate purification.

## Performing 96-Well Spin Plate Purification

### Recommended 96-Well Spin Plates

For large-scale procedures, you can use 96-well spin plates, such as the Gel Filtration Kit from Edge Biosystems.

**Note:** Other spin plate systems may be used to successfully remove unincorporated dye terminators. However, due to the large number of variables associated with using spin plate systems, you will need to optimize the performance of your system in your own laboratory.

### Purifying with the 96-Well Spin Plate

To perform purification with the spin plate:

1.	Prepare the extension products according to “Preparing Extension Products” on page 4-13, if desired.
2.	Prepare the 96-well spin plate per the manufacturer’s instructions.
3.	Perform the purification per the manufacturer’s instructions.  <b>IMPORTANT!</b> When using the Edge Biosystems gel filtration kit only, centrifuge at $850 \times g$ for 2 min.

# Selecting Sequencing Primers

---



## Selecting Sequencing Primers

**Overview** The choice of sequencing primer sequence, method of primer synthesis, and approach to primer purification can have a significant effect on the quality of the sequencing data obtained in dye terminator cycle sequencing reactions with this kit.

These decisions are particularly important when sequencing is done on real-time detection systems where signal strength is critical. Some of the recommendations given here are based on information that is general knowledge, while others are based on practical experience gained by Applied Biosystems scientists.

### Optimizing Primer Selection

The following recommendations are provided to help optimize primer selection:

- Primers should be at least 18 bases long to ensure good hybridization.
- Avoid runs of an identical nucleotide, especially guanine, where runs of four or more Gs should be avoided.
- Keep the G-C content in the range 30–80%.
- For cycle sequencing, primers with melting temperatures ( $T_m$ ) above 45 °C produce better results than primers with lower  $T_m$ .
- For primers with a G-C content less than 50%, it may be necessary to extend the primer sequence beyond 18 bases to keep the  $T_m > 45$  °C.
- Use of primers longer than 18 bases also minimizes the chance of having a secondary hybridization site on the target DNA.
- Avoid primers that have secondary structure or that can hybridize to form dimers.
- Several computer programs for primer selection are available. They can be useful in identifying potential secondary structure problems and determining if a secondary hybridization site exists on the target DNA.



# Control DNA Sequence

# B

## Control Sequence

**Partial Sequence of pGEM-3Zf(+)** The pGEM-3Zf(+) sequence below is the sequence of the –21 M13 forward primer, followed by the ensuing 1000 bases.

TGAAAAACGACGGCCAGT (–21 M13 primer)				
GAATTGTAAT	ACGACTCACT	ATAGGGCGAA	TTCGAGCTCG	40
GTACCCGGGG	ATCCTCTAGA	GTCGACCTGC	AGGCATGCAA	80
GCTTGAGTAT	TCTATAGTGT	CACCTAAATA	GCTTGCGTA	120
ATCATGGTCA	TAGCTGTTTC	CTGTGTGAAA	TTGTTATCCG	160
CTCACAATTC	CACACAACAT	ACGAGCCGGA	AGCATAAAGT	200
GTAAAGCCTG	GGGTGCCTAA	TGAGTGAGCT	AACTCACATT	240
AATTGCGTTG	CGCTCACTGC	CCGCTTTCCA	GTCGGGAAAC	280
CTGTCGTGCC	AGCTGCATTA	ATGAATCGGC	CAACGCGCGG	320
GGAGAGGCGG	TTTGCGTATT	GGGCGCTCTT	CCGCTTCCTC	360
GCTCACTGAC	TCGCTGCGCT	CGGTCGTTCCG	GCTGCGGCGA	400
GCGGTATCAG	CTCACTCAAA	GGCGGTAATA	CGGTTATCCA	440
CAGAATCAGG	GGATAACGCA	GGAAAGAACA	TGTGAGCAAA	480
AGGCCAGCAA	AAGGCCAGGA	ACCGTAAAAA	GGCCGCGTTG	520
CTGGCGTTTT	TCCATAGGCT	CCGCCCCCT	GACGAGCATC	560
ACAAAAATCG	ACGCTCAAGT	CAGAGGTGGC	GAAACCCGAC	600
AGGACTATAA	AGATACCAGG	CGTTTCCCCC	TGGAAGCTCC	640

CTCGTGCGCT	CTCCTGTTCC	GACCCTGCCG	CTTACCGGAT	680
ACCTGTCCGC	CTTTCTCCCT	TCGGGAAGCG	TGGCGCTTTC	720
TCATAGCTCA	CGCTGTAGGT	ATCTCAGTTC	GGTGTAGGTC	760
GTTCGCTCCA	AGCTGGGCTG	TGTGCACGAA	CCCCCGTTC	800
AGCCCGACCG	CTGCGCCTTA	TCCGGTAACT	ATCGTCTTGA	840
GTCCAACCCG	GTAAGACACG	ACTTATCGCC	ACTGGCAGCA	880
GCCACTGGTA	ACAGGATTAG	CAGAGCGAGG	TATGTAGGCG	920
GTGCTACAGA	GTTCTTGAAG	TGGTGGCCTA	ACTACGGCTA	960
CACTAGAAGG	ACAGTATTTG	GTATCTGCGC	TCTGCTGAAG	1000

# Sample Electrophoresis

---



## Some Important Reminders

- Dye set/primer (mobility) files for the BigDye<sup>®</sup> terminators v1.1 are the same as those for BigDye terminators original, but different than those for the dRhodamine terminators and BigDye terminators v3.1.
- If a mobility file for the wrong sequencing chemistry is used, some bases may be miscalled due to different dye labeling for the different chemistries.

**Note:** See “Required Software” on page 1-4 for information on obtaining the v1.1 dye set/primer (mobility) files.

# Electrophoresis on the ABI PRISM 3700 DNA Analyzer

## Requirements Run Modules

Configuration	Run Module
POP-5™ polymer, 50-cm	Seq1_1POP5DefaultModule
	Seq1_2POP5DefaultModule
POP-6™ polymer, 50-cm	Seq1_1POP6DefaultModule
	Seq1_2POP6DefaultModule

## Dye Set/Primer (Mobility) Files

Polymer	Dye Set/Primer (Mobility) File
POP-5 polymer	DT3700POP5{BD}v3.mob
POP-6 polymer	DT3700POP6{BD}v5.mob

## Standards

**IMPORTANT!** Use Dye Set E.

Dye Set	Standards
E	3700 BigDye® Terminator v1.1 Matrix Standard (PN 4336825)

**Note:** Refer to the product insert for instructions on using the standards for this instrument.

**Performing  
Sample  
Electrophoresis**

For information on how to perform sample electrophoresis on the 3700 instrument, refer to the following manuals:

- *ABI PRISM 3700 DNA Analyzer Sequencing Chemistry Guide* (PN 4309125)
- *ABI PRISM 3700 DNA Analyzer User's Manual* (PN 4306152)

# Electrophoresis on the ABI PRISM 3100 and 3100-Avant Genetic Analyzers

**Requirements** Electrophoresis and data analysis of samples on the ABI PRISM® 3100 and 3100-Avant Genetic Analyzers require the following:

## Run Modules

Configuration	Run Module
POP-4™ polymer, 36-cm	UltraSeq_POP4Default Module
POP-4 polymer, 80-cm	LongSeq80_POP4DefaultModule
POP-6™ polymer, 36-cm	RapidSeq36_POP6DefaultModule
POP-6 polymer, 50-cm	StdSeq50_POP6DefaultModule

## Dye Set/Primer (Mobility) Files

Polymer	Dye Set/Primer (Mobility) File
POP-4™ polymer	DT3100POP4LR{BD}v1.mob
POP-6™ polymer	DT3100POP6{BD}v1.mob

## Standards

**IMPORTANT!** Use Dye Set E.

Dye Set	Standards
E	3100 BigDye® Terminator v1.1 Matrix Standard (PN 4336824)

**Note:** Refer to the product insert for instructions on using the standards for this instrument.

**Performing  
Sample  
Electrophoresis**

For information on how to perform sample electrophoresis on the 3100 instrument, refer to the following manuals:

- *ABI PRISM 3100 Genetic Analyzer Sequencing Chemistry Guide* (PN 4315831)
- *ABI PRISM 3100 Genetic Analyzer User's Manual* (PN 4315834)

# Electrophoresis on the ABI PRISM 310 Genetic Analyzer

**Requirements** Electrophoresis and data analysis of samples on the ABI PRISM® 310 Genetic Analyzer requires the following:

## Filter Set E Run Modules

Configuration	Run Module
POP-4™ polymer, 1-mL syringe, 47-cm, 50-µm i.d. capillary, Ld = 36 cm	P4StdSeq (1 mL) E
POP-4 polymer, Rapid Sequencing, 1-mL syringe, 47-cm, 50-µm i.d. capillary, Ld = 36 cm	P4RapidSeq (1 mL) E
POP-6™ polymer, 1-mL syringe, 61-cm, 50-µm i.d. capillary	Seq POP6 (1 mL) E
POP-6 polymer, Rapid Sequencing, 1-mL syringe, 47-cm, 50-µm i.d. capillary	Seq POP6 Rapid (1 mL) E

## Dye Set/Primer (Mobility) Files

Polymer	Operating System	Dye Set/Primer (Mobility) File
POP-4™ polymer	NT®	DT310POP4{BD}v2.mob
POP-6™ polymer		DT310POP6{BD-LR}v3.mob DT310POP6{BD}.mob
POP-4™ polymer	Macintosh®	DT310POP4{BD}v2.mob
POP-6™ polymer		DT POP6{BD Set-AnyPrimer}

## Standards

**IMPORTANT!** The instrument (matrix) file for the BigDye terminators v1.1 cannot be used for the BigDye terminators v3.0 or v3.1.

Dye/Filter Set	Standards for Instrument (Matrix) File Generation
E	310/377 BigDye Terminator v1.1 Matrix Standards (PN 4336805)

**Note:** Refer to the product insert for instructions on using the standards for this instrument.

## Performing Sample Electrophoresis

For information on how to perform sample electrophoresis on the 310 instrument, refer to the following manuals:

- *ABI PRISM 310 Genetic Analyzer Sequencing Chemistry Guide* (PN 4303189)
- *ABI PRISM 310 Genetic Analyzer User's Manual* (PN 4317588)

# Electrophoresis on the ABI PRISM 377 DNA Sequencers

**Requirements** Electrophoresis and data analysis of samples on the ABI PRISM® 377 DNA Sequencers (all models\*) require the following:

## Filter Set E Run Modules

Configuration	Run Module
36-cm wtr, 1200 scans/hr, any comb	Seq Run 36E-1200
36-cm wtr, 2400 scans/hr, any comb	Seq Run 36E-2400
48-cm wtr, 1200 scans/hr, any comb	Seq Run 48E-1200

\*Any plate check and prerun module can be used on the ABI PRISM 377 DNA Sequencers.

## Dye Set/Primer (Mobility) Files

Gel Formulation	Operating System	Dye Set/Primer (Mobility) File
4.5% acrylamide (29:1) or 5% Long Ranger™ gel	NT	DT377{BD}.mob
	Macintosh	DT{BD Set Any-Primer}

\*Includes the ABI PRISM 377, ABI PRISM 377-18, ABI PRISM 377 with XL Upgrade, and the ABI PRISM 377 with 96-Lane Upgrade instruments.

## Standards

**IMPORTANT!** The instrument (matrix) file for the BigDye terminators v1.1 cannot be used for the BigDye terminators v3.0 or v3.1.

Dye/Filter Set	Standards for Instrument (Matrix) File Generation
E	310/377 BigDye Terminator v1.1 Matrix Standards (PN 4336805)

**Note:** Refer to the product insert for instructions on using the standards for this instrument.

### Using the Lane Guide Kit

If you are using the BigDye® chemistries v1.1 on the 377 instrument in conjunction with the ABI PRISM® Lane Guide™ Lane Identification Kit, refer to that kit's protocol (PN 4313804) for instructions on resuspending and loading samples.

### Using Long-Read Gel and Buffer Formulations

For longer sequencing read lengths follow the gel and buffer formulations described in the user bulletin entitled *Achieving Longer High Accuracy Reads on the 377 Sequencer* (PN 4315153).

### Performing Sample Electrophoresis

For information on how to perform sample electrophoresis on the 377 instrument, refer to the following manuals:

- *Automated DNA Sequencing Chemistry Guide* (PN 4305080)
- *ABI PRISM 377 DNA Sequencer User's Manual* (PN 4307164)



# Troubleshooting

# D

Observation	Possible Causes	Recommended Actions
No recognizable sequence	Insufficient template	<ul style="list-style-type: none"><li>• Quantitate DNA template</li><li>• Increase amount of DNA in the sequencing reactions</li></ul>
	Inhibitory contaminant in the template	Clean up the template
	Insufficient primer	<ul style="list-style-type: none"><li>• Quantitate the primer</li><li>• Increase amount of primer in the sequencing reactions</li></ul>
	Primer has no annealing site	Use a primer that is complementary to the template
	Poor primer design or incorrect primer sequence	Redesign the primer
	Missing reagent	Repeat the reactions, carefully following the protocol
	Old or mishandled reagents	Use fresh reagents
	Thermal cycler power failure	Repeat the reactions
	Thermal cycling conditions incorrect	<ul style="list-style-type: none"><li>• Calibrate the thermal cycler regularly</li><li>• Use correct thermal cycling parameters</li><li>• Use correct tube for your thermal cycler</li><li>• Set ramp rate to 1 °C/sec</li></ul>
Extension products lost during reaction cleanup	<ul style="list-style-type: none"><li>• Make sure you use the correct centrifugation speeds and times for precipitation and spin column procedures</li><li>• Make sure the ethanol concentration is correct for the precipitation protocols</li></ul>	

<b>Observation</b>	<b>Possible Causes</b>	<b>Recommended Actions</b>
No recognizable sequence (continued)	Extension products not resuspended	Carefully resuspend the sample pellet in loading buffer
	Lane tracking failure (377 instrument)	Check the lane tracking. If necessary, retrack and reextract the lanes
	Electrokinetic injection failure (capillary instruments)	Repeat the injections
Noisy data throughout the sequence with low signal strength	Insufficient DNA in the sequencing reactions	<ul style="list-style-type: none"> <li>• Use more DNA in the sequencing reactions</li> <li>• Load or inject more of the resuspended sequencing reactions</li> </ul>
	Degraded template	Prepare fresh DNA and repeat the reactions
	Old or mishandled reagents	Use fresh reagents
	Thermal cycling conditions incorrect	<ul style="list-style-type: none"> <li>• Calibrate the thermal cycler regularly</li> <li>• Use correct thermal cycling parameters</li> <li>• Use correct tube for your thermal cycler</li> <li>• Set ramp rate to 1 °C/sec</li> </ul>
	Lane tracking failure (377 instrument)	Check the lane tracking. If necessary, retrack and reextract the lanes
	Electrokinetic injection failure (capillary instruments)	Repeat the injections
Noisy data throughout the sequence with good signal strength	Inhibitory contaminant in the template	Clean up the template
	Multiple templates in the sequencing reaction	Examine template on an agarose gel to be sure only one template is present.
	Multiple priming sites	<ul style="list-style-type: none"> <li>• Make sure primer has only one priming site.</li> <li>• If necessary, redesign primer</li> </ul>
	Multiple primers when sequencing PCR products	Purify your PCR template to remove excess primers

<b>Observation</b>	<b>Possible Causes</b>	<b>Recommended Actions</b>
Noisy data throughout the sequence with good signal strength (continued)	Primer with N-1 contamination	Use HPLC-purified primer
	High signal saturating the detector	<ul style="list-style-type: none"> <li>• Use less DNA in the sequencing reactions</li> <li>• Load or inject less of the resuspended sequencing reactions</li> </ul>
	Incorrect run module	Use correct run module
	Incorrect instrument (matrix) file	Use correct instrument file for terminator chemistry
Noise up to or after a specific point in the sequence	Mixed plasmid separation	Make sure you have only one template
	Multiple PCR products	Make sure you have only one template
	Primer-dimer contamination in PCR sequencing	<ul style="list-style-type: none"> <li>• Optimize your PCR amplification</li> <li>• Make sure there is no sequence complementarity between the two PCR primers</li> <li>• Make sure your sequencing primer does not overlap the sequences of the PCR primers</li> <li>• Use a Hot Start technique such as with Amplitaq® Gold® polymerase</li> </ul>
	Slippage after repeat region in template	<ul style="list-style-type: none"> <li>• Try an alternate sequencing chemistry</li> <li>• Use an anchored primer</li> </ul>
Poor mobility correction	Incorrect dye set/primer (mobility) file	Use correct dye set/primer file
	Incorrect Peak 1 location for data analysis	Select a new Peak 1 location
	Gel with very different separation properties than the gel matrices used to construct the dye set/primer (mobility) file	Use correct dye set/primer file for your gel type

Observation	Possible Causes	Recommended Actions
<p>Excess dye peaks at beginning of sequence</p>	<p>Poor removal of unincorporated dye terminators</p>	<p>Use ethanol/EDTA precipitation protocol to remove unincorporated dye terminators</p>
		<p>With Centri-Sep spin columns, take care to load sample on the center of the gel surface</p> <p><b>Note:</b> Do not touch the gel surface with the pipet tip.</p> <p><b>IMPORTANT!</b> Be sure you hydrate the column for at least 2 hours.</p>
		<p>Spin samples for recommended times (spinning too long precipitates more dyes with the sample)</p>
		<p>With microfuge tubes, aspirate the supernatant rather than decanting (decanting leaves excess ethanol on the sides of the tube)</p>
		<p>Select the start point for data analysis to exclude excess dye peaks</p>
<p>Pull-up peaks or bleedthrough</p>	<p>Total signal strength over 4000</p>	<ul style="list-style-type: none"> <li>• Quantitate DNA template</li> <li>• Decrease amount of DNA in the sequencing reactions</li> </ul>
		<p>Load or inject less of the resuspended sequencing reactions</p>

# Obtaining Technical Support

---



## Services and Support

### **Applied Biosystems Web Site**

A services and support page is available on the Applied Biosystems Web site. To access this, go to:

**<http://www.appliedbiosystems.com>**

and click the link for services and support.

At the services and support page, you can:

- Search through frequently asked questions (FAQs)
- Submit a question directly to Technical Support
- Order Applied Biosystems user documents, MSDSs, certificates of analysis, and other related documents
- Download PDF documents
- Obtain information about customer training
- Download software updates and patches

In addition, the services and support page provides worldwide telephone and fax numbers to contact Applied Biosystems Technical Support and Sales facilities.



# Index

## Numerics

- 310 requirements
    - mobility files C-6
    - run modules C-6
    - sequencing standards C-7
  - 3100 requirements
    - mobility files C-4
    - run modules C-4
    - spectral calibration standards C-4
  - 3700 requirements
    - mobility files C-2
    - run modules C-2
    - spectral calibration standards C-2
  - 377 requirements
    - matrix file standards C-9
    - mobility files C-8
    - run modules C-8
- ## A
- attention words 1-11
- ## B
- bacterial artificial chromosomes 2-3
- ## C
- centrifuge, recommended 1-9
  - components, sequencing kits 1-6
  - contaminants, of templates 2-5
  - control
    - double-stranded 2-1
    - electrophoresis 2-1
    - single-stranded 2-1
  - control DNA, reason for 2-1
  - customer support. *See* technical support E-1

- cycle sequencing
  - BAC DNA 3-7
  - protocol for GeneAmp 3-5

## D

- disposables
  - cycle sequencing 1-8
  - purifying 1-9
- DNA
  - quantity characteristics 2-6
  - quantity to use 2-6
- dye set/primer files, downloading 1-6
- dye/filter set
  - 3100 1-4
  - 3700 1-4
  - 377 1-4

## E

- ethanol/EDTA precipitation
  - 384-well 4-4
  - 96-well 4-3
- ethanol/EDTA/sodium acetate
  - precipitation in 384-well 4-8
  - precipitation in 96-well 4-6

## F

- formulation, new 1-1
- freeze-thaw, limits 1-7

## H

- hazards
  - chemical 1-11
  - waste 1-12

---

## I

instruments  
sequencing 1-3  
thermal cyclers 1-3

## L

Lane Guide C-9  
long read lengths C-9

## M

matrix standards  
310 1-4  
377 1-4  
general requirements 1-2  
mobility files 1-2  
about C-1  
downloading 1-6  
MSDSs, ordering 1-13

## P

pGEM-3Zf(+) sequence B-1  
primers  
importance of selection A-1  
optimizing selection A-1  
purification  
SDS and 96-well 4-13  
spin plate 4-14  
purifying  
methods 2-3, 4-1  
methods for PCR 2-3  
PCR by ultrafiltration 2-4

## Q

quality problems, template 2-5  
quality, examining DNA 2-5  
quantitating DNA 2-6

## R

reaction mixtures  
BAC DNA 3-7  
for 384-well 3-4  
for 96-well 3-2  
for microcentrifuge tubes 3-2  
reagents  
electrophoresis 1-10  
purifying 1-9  
required files  
matrix 1-1  
spectral 1-1

## S

safety  
chemical 1-11 to 1-13  
guide 1-11 to 1-13  
SDS, and purification 4-13  
sequencing buffer  
optimizing 1-2  
sequencing kits, ordering 1-6  
sequencing problems 2-4  
sequencing standards  
for each instrument 2-2  
general requirements 1-2  
shorter read lengths, and DNA quantity 2-6  
signal intensities, and DNA quantity 2-6  
site preparation and safety guide 1-11  
spectral standards  
3100 1-4  
3700 1-4  
spin column purification  
optimizing 4-10  
performing 4-11  
storage temperatures 1-7

---

## T

technical support E-1

templates

    BAC DNA 2-3

    PCR 2-3

    poor quality characteristics 2-4

    preparing 2-3

top-heavy data, and DNA quantity 2-6

## U

user attention words 1-11

## W

warning symbols 1-11





**Headquarters**

850 Lincoln Centre Drive  
Foster City, CA 94404 USA  
Phone: +1 650.638.5800  
Toll Free (In North America): +1 800.345.5224  
Fax: +1 650.638.5884

**Worldwide Sales and Support**

Applied Biosystems vast distribution and service network, composed of highly trained support and applications personnel, reaches 150 countries on six continents. For sales office locations and technical support, please call our local office or refer to our Web site at [www.appliedbiosystems.com](http://www.appliedbiosystems.com).

**[www.appliedbiosystems.com](http://www.appliedbiosystems.com)**



Applera Corporation is committed to providing the world's leading technology and information for life scientists. Applera Corporation consists of the Applied Biosystems and Celera Genomics businesses.

Printed in USA, 09/2002  
Part Number 4337036 Rev. A

an **Applera** business